Autocontouring and Manual Contouring: Which Is the Better Method for Target Delineation Using ¹⁸F-FDG PET/CT in Non–Small Cell Lung Cancer?

Kailiang Wu¹, Yee C. Ung², David Hwang³, Ming S. Tsao³, Gail Darling⁴, Donna E. Maziak⁵, Romeo Tirona², Kathy Mah², and C. Shun Wong²

¹Department of Radiotherapy, Shanghai Cancer Center, Fudan University, Shanghai, China; ²Odette Cancer Centre, Sunnybrook Health Sciences Centre, University of Toronto, Toronto, Ontario, Canada; ³Department of Pathology, Princess Margaret Hospital, University of Toronto, Toronto, Ontario, Canada; ⁴Department of Thoracic Surgery, Princess Margaret Hospital, University of Toronto, Toronto, Ontario, Canada; and ⁵Ottawa Regional Cancer Center, University of Ottawa, Ottawa, Canada

Previously, we showed that a CT window and level setting of 1.600 and -300 Hounsfield units, respectively, and autocontouring using an ¹⁸F-FDG PET 50% intensity level correlated best with pathologic results. The aim of this study was to compare this autocontouring with manual contouring, to determine which method is better. Methods: Seventeen patients with non-small cell lung cancer underwent ¹⁸F-FDG PET/CT before surgery. The maximum diameter on pathologic examination was determined. Seven sets of gross tumor volumes (GTVs) were defined. The first set (GTV_{CT}) was contoured manually using only CT information. The second set (GTV_{Auto}) was autocontoured using a 50% intensity level for ¹⁸F-FDG PET images. The third set (GTV_{Manual}) was manually contoured using a visual method on PET images. The other 4 sets combined CT and ¹⁸F-FDG PET images fused to one another to become composite volumes: $GTV_{CT+Auto}$, $GTV_{CT+Manual}$, $GTV_{CT-Auto}$, and $GTV_{CT-Manual}$. To quantitate the degree to which CT and ¹⁸F-FDG PET defined the same region of interest, a matching index was calculated for each case. The maximum diameter of GTV was compared with the maximum diameter on pathologic examination. Results: The median GTV_{CT}, GTV_{Auto}, GTV_{Manual}, $\text{GTV}_{\text{CT}+\text{Auto}}, \text{GTV}_{\text{CT}+\text{Manual}}, \text{GTV}_{\text{CT}-\text{Auto}}, \text{and } \text{GTV}_{\text{CT}-\text{Manual}}$ were 6.96, 2.42, 4.37, 7.46, 10.17, 2.21, and 3.38 cm³, respectively. The median matching indexes of GTV_{CT} versus GTV_{CT+Auto}, GTV_{Auto} versus $GTV_{CT+Auto}$, GTV_{CT} versus $GTV_{CT+Manual}$, and GTV_{Manual} versus $GTV_{CT+Manual}$ were 0.86, 0.65, 0.88, and 0.81, respectively. Compared with the maximum diameter on pathologic examination, the correlations of GTV_{CT}, GTV_{Auto}, $\text{GTV}_{\text{Manual}},\ \text{GTV}_{\text{CT+Auto}},\ \text{and}\ \text{GTV}_{\text{CT+Manual}}$ were 0.87, 0.83, 0.93, 0.86, and 0.94, respectively. Conclusion: The matching index was higher for manual contouring than for autocontouring using a 50% intensity level on ¹⁸F-FDG PET images. When using a 50% intensity level to contour the target of non-small cell lung cancer, one should also consider using manual contouring of ¹⁸F-FDG PET to check for any missed disease.

For correspondence or reprints contact: Yee C. Ung, Sunnybrook Odette Cancer Centre, 2075 Bayview Ave., Toronto, Ontario, Canada M4N 3M5.

E-mail: yee.ung@sunnybrook.ca

Key Words: gross tumor volume; target definition; FDG PET/ CT; pathology; non-small cell lung cancer

J Nucl Med 2010; 51:1517–1523 DOI: 10.2967/jnumed.110.077974

t is well established that ¹⁸F-FDG PET plays an important role in the staging of non–small cell lung cancer (NSCLC). Multiple studies have demonstrated the utility of PET for improving staging accuracy. In an overview of the available literature, ¹⁸F-FDG PET was found to have a 79%–100% sensitivity and a 40%–90% specificity in diagnosing primary lung cancer (1).

The conventional imaging modality for therapy planning in NSCLC is CT. Targeting of the gross tumor has been facilitated by the use of CT simulation, which allows for more accurate delineation of the tumor. Multimodality imaging combining functional and anatomic information such as PET has allowed for further refinement in the therapy planning process and has a significant impact on the planning target volume. However, much uncertainty exists regarding the most appropriate threshold cutoff that should be used to define a PET target volume in NSCLC therapy planning. There are sparse data on gross tumor volume (GTV) contoured differences using different PET/CT thresholds correlated with gross tumor size on pathologic examination. Different institutions use varying methods to define the PET volume; these include the halo phenomenon, the absolute standardized uptake value (SUV) (2), a regressive SUV function threshold (3), a percentage of the maximum SUV intensity levels (4-7), an identification of affected anatomic structures by ¹⁸F-FDG (8), and a simply visual evaluation of PET images (9-12). These methods have demonstrated huge alterations in the target volume between CT-based therapy planning and PET/CT-based therapy planning. Because of a lack of uniformity in defining the PET tumor contours in the published literature, interpretation of the available data is difficult and leaves

inm077974-sn ■ 9/9/10 Copyright 2010 by Society of Nuclear Medicine.

Received Apr. 7, 2010; revision accepted Jul. 20, 2010.

COPYRIGHT © 2010 by the Society of Nuclear Medicine, Inc.

clinicians uncertain as to how they should incorporate PET into the therapy planning process. Because of possible incomplete tumor coverage, the 40%-of-maximum-SUV concept did not appear generally suitable for target volume delineation, and the contrast-oriented methods for contour definition showed more satisfactory results (13). Our previous study (14) examined the impact of varying the CT window and level parameters and the PET intensity thresholds on the radiologic tumor volumes as compared with the measured diameter on pathologic examination. Our results showed that setting a 50% intensity level on PET and a CT window of 1,600 Hounsfield units (HU) and level of -300HU correlated best with the maximum diameter of the primary tumor as measured on pathologic examination.

The aim of this study was to investigate the difference in GTV autocontouring using a PET 50% intensity level and using manual contouring by a visual method, compared with pathologic results, and to determine which is the better method for target delineation on PET/CT images.

MATERIALS AND METHODS

Characteristics of Patients and Tumors

This study was approved by our local institutional research ethics board. Each patient was required to have pathologic confirmation of NSCLC. Seventeen patients with surgically resectable NSCLC underwent PET/CT before surgery between December 2004 and May 2007. The characteristics of the patients [**Table 1**] and primary tumors are shown in Table 1.

PET/CT Acquisition and Image Registration

Patients were asked to consume a high-protein, low-carbohydrate diet (to reduce myocardial uptake of ¹⁸F-FDG) and to avoid vigorous exercise in the 24 h preceding imaging. Patients fasted at least 6 h before the intravenous injection of ¹⁸F-FDG. Blood glucose levels were checked and recorded. The amount of ¹⁸F-FDG injected was 185–370 MBq, depending on patient weight. Patients rested for approximately 1 h before imaging. Freebreathing PET and CT images were acquired using a Gemini PET/CT scanner (Philips) or a Discovery PET/CT scanner (GE Healthcare). The interval between ¹⁸F-FDG injection and the start of the PET acquisition was 40–60 min. First, a topogram was made from the skull to the mid thigh. Second, CT images (3-mm slices) at an interval of 3 mm typically were obtained from the base of the skull through the proximal thighs without the administration of either oral or intravenous contrast agents.

PET data were acquired using an acquisition time of 3 min per table position with a 50% overlap. The data were reconstructed using a 3-dimensional row-action maximum-likelihood algorithm for the Gemini and ordered-subsets expectation maximization for the Discovery and corrected for attenuation using a CT-derived transmission map. The voxel size after reconstruction was $4.0 \times 4.0 \times 4.0$ mm.

Once the PET and CT images were acquired, the image datasets were transferred to a Pinnacle therapy planning workstation (version 8.0, Philips) for image coregistration.

GTV Definition and Delineation

Target volume was defined according to the prepared protocol before the study. Briefly, gross macroscopic tumor visible on the CT or PET scans defined the GTV. Only primary lung tumors were contoured. Guidelines were established for contouring GTVs. A standardized window-level setting was used to manually contour the GTV on PET images.

Seven sets of GTVs were defined by a physician observer (Table 2; Fig. 1), with agreement by a second physician observer. [Table 2] GTV_{CT} was delineated manually using only CT information. [Fig. 1] GTV_{Auto} was autocontoured using a 50% intensity level on PET images. GTV_{Manual} was manually contoured using a visual method only on PET images. The other 4 GTVs combined CT and PET images fused to one another to become composite volumes: GTV_{CT+Auto}, GTV_{CT+Manual}, GTV_{CT-Auto}, and GTV_{CT-Manual}, where plus (+) means the volume must include both GTV_{CT} and GTV Auto or GTV Manual and minus (-) means the volume includes the intersection of both GTV_{CT} and GTV_{Auto} or GTV Manual (Figs. 1C and 1D). This volume was a composite delineated by initially creating a union of the CT and PET regions of interest. For example, when GTV_{CT+Manual} was contoured, the first GTV contour was obtained using CT information only, and then, using a visual method, another GTV contour was obtained using PET information. Finally, the last manual contour was a composite GTV_{CT+Manual} using PET/CT fusion. When the GTV was contoured using the PET image, the maximum intensity was measured within the primary tumor using voxel values. When the GTV_{Auto} was contoured, the planning software automatically localized the PET-imaged object using a threshold value of 50% intensity relative to the maximum intensity within the primary tumor on each PET transaxial image. Using PET, GTV_{Manual}

TABLE 1. Characteristics of Patients and Primary Tumors $(n = 17)$			
Characteristic	n	Median	Range
Age (y)		59	50–85
Sex			
Male	9		
Female	8		
LUL	7		
LLL	2		
RUL	6		
RLL	2		
Maximum diameter of		2.1	1.2–6.5
primary tumor on			
<3 cm	11		
3–5 cm	5		
>5 cm	1		
Histologic type			
Squamous cell carcinoma	3		
Adenocarcinoma	13		
Histologic grade	1		
1	3		
2	8		
3	4		
No specific grade	2		
Maximum SUV intensity level in primary tumor	17	8,221	2,318–38,686
LUL = left upper lobe: LLL	_ = lef	t lower lo	be; RUL = riaht
upper lobe; RLL = right lower	lobe.		

1518 THE JOURNAL OF NUCLEAR MEDICINE • Vol. 51 • No. 10 • October 2010

TABLE 2. Definition of GTVs			
GTV	Contoured image	Included volume	
GTV _{CT}	СТ	Volume by manual contouring on CT image	
GTV _{Auto}	PET	Volume by autocontouring on PET image	
GTV _{Manual}	PET	Volume by manual contouring on PET image	
GTV _{CT+Auto}	PET/CT	Volume including both GTV _{CT} and GTV _{Auto}	
GTV _{CT+Manual}	PET/CT	Volume including both GTV _{CT} and GTV _{Manual}	
GTV _{CT-Auto}	PET/CT	Volume including only intersection volume between GTV_{CT} and GTV_{Auto}	
GTV _{CT-Manual}	PET/CT	Volume including only intersection volume between $\mathrm{GTV}_{\mathrm{CT}}$ and $\mathrm{GTV}_{\mathrm{Manual}}$	

was defined as the PET-visualized enhancing gross tumor (2), with the GTV edge positioned at the maximum local gradient magnitude distinguishable by the observer.

GTV_{CT} was manually delineated using a lung window of 1,600 HU and a level of -300 HU on each transaxial image without knowledge of the PET results. GTV_{CT} always was contoured first. When the observer was drawing the GTV_{CT} contours, the PET image was masked. All areas of gross primary lung tumor were contoured using lung window settings for the interface between tumor and lung and using modified mediastinal window settings (window, 600 HU; level, 40 HU) for the interface if the tumor was close to the mediastinum.

The maximum diameter of GTV was measured on each PET or CT slice in transverse, anteroposterior, and craniocaudal directions. The greatest of these diameters was used for comparison with the maximum diameter on pathologic examination. The total GTV contoured on the CT or PET image or composite GTV was measured and calculated for each patient in the 3-dimensional reconstruction images by planning software.

GTV Comparison

The GTVs as defined by the CT scans alone, by the PET scans using autocontouring or manual contouring, and by the fused PET/ CT scans were compared. To quantitate the degree to which CT and PET (autocontouring or manual contouring) defined the same region of interest, we calculated a matching index using previously described methods. In brief, a matching index was calculated for each case using the following formula: $\{[(a - a/b) + (b - b/a)]/a + b\}$ (15).

Pathology Procedure

All 17 patients underwent surgical resection of their lung tumor. Immediately after surgery, the involved lung lobes were inflated and fixed for 12–24 h in 10% neutral buffered formalin. Specimens were then serially sectioned at 3- to 5-mm intervals, and the maximum diameter of the primary tumor was measured by macroscopic examination in 3 dimensions. Sections for histologic examination of the tumor and lymph nodes were taken and processed for paraffin embedding and sectioning using standard protocols.

Statistical Analysis

GraphPad Prism statistical analysis software was used. Oneway ANOVA was used to determine differences among groups. Concerning the delineations, the different volumes were measured for each case. Results are summarized by the mean, median, and range. The mean difference among the GTVs was calculated for tumors at different contouring methods. Linear and logarithmic regression analyses were used to determine the relationship



FIGURE 1. CT (A) and PET (B) images showing GTVs. Pink, green, lavender, and yellow contours represent GTV_{CT} , GTV_{Auto} , $GTV_{CT+Auto}$, and $GTV_{CT-Auto}$, respectively. GTV_{CT} is delineated manually using only CT information. GTV_{Auto} is autocontoured using 50% intensity level on PET images. $GTV_{CT-Auto}$ is volume of intersection of both GTV_{CT} and GTV_{Auto} (C, solid line). $GTV_{CT+Auto}$ must include both GTV_{CT} and GTV_{Auto} (D, solid line).

RGB

TABLE 3. GTV for Autocontouring or Manual Contouring				
GTV	Mean (cm ³)	Median (cm ³)	Range (cm³)	P*
GTV _{CT}	21.26	6.96	2.14–114.28	>0.05
GTV _{Auto}	7.19	2.42	0.67–31.08	>0.05
GTV _{Manual}	23.02	4.37	1.52-130.83	>0.05
GTV _{CT+Auto}	23.72	7.46	2.87-115.36	>0.05
GTV _{CT+Manual}	16.70	10.17	3.31–103.47	>0.05
GTV _{CT-Auto}	7.74	2.21	0.69–39.74	>0.05
GTV _{CT-Manual}	13.65	3.38	0.63–94.60	>0.05
*1-way ANOV	Ά.			

between the maximum diameter of GTV and the pathologic maximum diameter. The Pearson correlation coefficient was used to compare pathologic and imaging estimates of maximum tumor diameter. Two-tailed P values are provided, and P values of less than 0.05 were considered statistically significant.

RESULTS

GTVs

[Table 3] The GTVs are shown in Table 3. GTV_{CT} was larger than GTV_{Auto} in 88% of cases (15/17), GTV_{Auto} was larger than GTV_{CT} in 12% (2/17), GTV_{CT} was larger than GTV_{Manual} in 76% (13/17), and GTV_{Manual} was larger than GTV_{CT} in 24% (4/17).

Autocontouring and Manual Contouring Matching Index in Composite Volume

The matching index of GTV_{CT} , GTV_{Auto} , and GTV_{Manual} [**Table 4**] versus composite GTVs are shown in Table 4. The median matching index of GTV_{CT} versus $GTV_{CT+Auto}$, GTV_{Auto} versus $GTV_{CT+Auto}$, GTV_{CT} versus $GTV_{CT+Manual}$, and GTV_{Manual} versus $GTV_{CT+Manual}$ was 0.86, 0.65, 0.88, and 0.81, respectively. The matching index was higher with PET manual contouring than with PET autocontouring. However, there was no significant difference between groups except for GTV_{CT} versus $GTV_{CT+Auto}$ compared with GTV_{CT} versus $GTV_{CT-Auto}$.

Radiologic–Pathologic Correlation

A strong correlation was found between maximum diameter on pathologic examination and maximum diameter

of GTVs (Pearson correlation coefficient, Table 5). The best [**Table 5**] correlation with pathology was found with $\text{GTV}_{\text{CT+Manual}}$ and $\text{GTV}_{\text{Manual}}$. $\text{GTV}_{\text{CT+Manual}}$ correlated better with pathology than did $\text{GTV}_{\text{CT+Auto}}$ (Fig. 2). Also, $\text{GTV}_{\text{Manual}}$ correlated [Fig. 2] better with pathology than did GTV_{Auto} .

DISCUSSION

PET is a significant advance in cancer imaging with great potential for optimizing radiation therapy planning (16). The delineation of target volume is a critical step in high-precision radiation therapy planning (17). Both good image quality and good delineation protocols are crucial for target volume definition (18). In this study, we evaluated the degree to which PET autocontouring and manual contouring defined the same volume of interest in NSCLC.

It is clear that with PET-defined tumor volumes in radiation therapy planning, variations in setting the image signal thresholds can significantly affect the contour of the GTV. Various methods are currently used to determine the outline of ¹⁸F-FDG-positive tissue. There is no validated standardized method for setting this threshold. The simplest method, which is widely used, is visual interpretation of the PET scan and definition of contours as judged visually in cooperation with an experienced nuclear medicine physician (9,19,20). Another method is the SUV, using percentages of the maximum SUV and regression function or source-to-background ratio (3). Published methods based on a threshold determined as a percentage of the maximum SUV (percentage threshold) have used values ranging from 15% to 50% (2,4,8,9,14,19,21–25). The reported variability of threshold values for lung lesions of different volumes indicates that there is no standard value applicable to all patients and that techniques for setting individual thresholds need to be defined and standardized. Although many have used a percentage of the maximum SUV intensity to define a tumor on PET, some would suggest that this fixedthreshold intensity is inadequate for target volume definition and tends to underestimate target volumes (3). Despite these views, we believe that there are several advantages to using a fixed-threshold intensity, as shown in our previous study (14) in which only 70% of patients could be contoured using a 50% intensity level.

TABLE 4. Matching Index in Composite GTV					
Parameter	Mean ± SD	Median	Range	Р	
GTV _{CT} vs. GTV _{CT+Auto}	0.84 ± 0.14	0.86	0.59-0.99	<0.05*	
GTV _{Auto} vs. GTV _{CT+Auto}	0.63 ± 0.31	0.65	0.08–0.97	>0.05	
GTV _{CT} vs. GTV _{CT+Manual}	0.84 ± 0.13	0.88	0.59–0.99	>0.05	
GTV _{Manual} vs. GTV _{CT+Manual}	0.80 ± 0.15	0.81	0.45–0.99	>0.05	
GTV _{CT} vs. GTV _{CT-Auto}	0.59 ± 0.34	0.62	0.04–0.98	>0.05	
GTV _{Auto} vs. GTV _{CT-Auto}	0.58 ± 0.35	0.70	0.03–0.97	>0.05	
GTV _{CT} vs. GTV _{CT-Manual}	0.73 ± 0.22	0.75	0.24–0.99	>0.05	
GTV _{Manual} vs. GTV _{CT-Manual}	0.73 ± 0.24	0.72	0.20–0.99	>0.05	

*1-way ANOVA for matching index of GTV_{CT} vs. $\text{GTV}_{\text{CT+Auto}}$ compared with GTV_{CT} vs. $\text{GTV}_{\text{CT-Auto}}$.

TABLE 5. Correlation Coefficient of Maximum Diameter of GTV Compared with Maximum Diameter of Tumor on Pathologic Examination

Parameter	Median (cm)	Range (cm)	Р	Pearson r	
Pathologic	2.10	1.20-6.50			
GTV _{CT}	3.24	2.10-8.84	< 0.0001	0.87	
GTV _{Auto}	3.08	1.73–7.41	< 0.0001	0.83	
GTV_{Manual}	3.11	2.19-8.91	< 0.0001	0.93	
$GTV_{CT+Auto}$	3.51	2.31-8.92	< 0.0001	0.86	
GTV _{CT+Manual}	3.66	2.07-8.67	< 0.0001	0.94	

The gold standard for validating a threshold technique for tumor definition would be a comparison with histologic specimens. This poses a particular problem in lung cancer, in which accurate spatial correlation of excised surgical specimens with imaging is difficult to achieve. As the use of PET increases in controlled studies reporting tumor outcomes or correlating PET volumes with pathologic data, we may have enough information to develop a more unified definition for PET volume contouring. Our previous study compared the maximum diameter of the tumor on pathologic examination with the maximum diameter of the GTV in 3-dimensional contouring on PET or CT images. Our results show that the 50% intensity level and a CT window of 1,600 HU and level of -300 HU correlated best to the actual geometry (14). In the current study, the matching index was higher with PET manual contouring than with PET autocontouring. However, we only correlated the measured maximum diameter on pathology.

Many factors affect SUV measurements and, therefore, tumor contours (e.g., the metabolic activity of a tumor, heterogeneity within a tumor, and tumor motion), thus making SUV-based contouring methods unreliable. Phantom studies (10) can help define thresholds for percentage maximum uptake in defining the tumor edge but have the limitation of not being able to adequately simulate the effect of background uptake as seen clinically (8,21,23,25,26). ¹⁸F-FDG PET–pathology correlation could be useful for building on existing data (27). Such studies could help characterize tumor boundaries, assist in image segmentation, and aid in the understanding of motion and apparent

tumor volume. It is important to remember that a PET image is typically an average of multiple respiratory cycles and that a standard free-breathing CT image is simply a snapshot in time. Investigations from Memorial Sloan-Kettering Cancer Center have found that the variation in maximum SUV may be as great as 24% between end inspiration and end expiration or significant tumor size reductions (28). It is possible that PET tumor-volume contouring methods that use fixed-intensity thresholds are therefore more susceptible to motion error. Inaccurate PET/CT fusion methods could also be problematic. Capitalizing on digital wholemount histopathologic methods developed by Clarke et al. (29), the aim is to generate 3-dimensional pathologic reconstructions and then coregister and compare these with ¹⁸F-FDG PET/CT volumes. Although significant early challenges have been encountered in working with lung tissue, such studies are in their infancy. Stroom et al. (30) have also tried this work in NSCLC. They investigated the feasibility of pathology-correlated imaging for lung tumors, taking into account lung deformations after surgery. Their results have shown that pathology-correlated lung imaging is feasible and can be used to improve target definition. Geets et al. (31) have addressed similar questions in head and neck cancer, finding the ¹⁸F-FDG PET GTV more reflective of pathology than is the MRI or CT GTV. van Baardwijk et al. (18) also observed a good correlation between the maximum diameter on PET scans using an autocontouring method based on source-to-background ratio and using the macroscopic tumor diameter of the surgical specimens from 23 cases of operable lung cancer.

None of the methods of automatic delineation of 18 F-FDG–positive tissue can be regarded as reliable or standard. Because the published delineation method greatly influences the size and shape of the GTV, in multicenter studies detailed protocols that include phantom evaluation and quality control must be followed (*16*). Until now, no objective data have been available on this point. The best method to date, as seen by many, is for the radiooncologist to use contouring. Contouring requires an experienced nuclear medicine specialist following a predefined protocol (*16*). A significant limitation of our study is the potential observer bias introduced by the use



FIGURE **2.** (A) Linear rearession between maximum diameter of GTV_{CT+Auto} and maximum diameter of primary tumor (r = 0.86). (B) Linear regression between maximum diameter of $\text{GTV}_{\text{CT}+\text{Manual}}$ and maximum diameter of primary tumor (r = 0.94). Maximum diameter of primary tumor was determined on macroscopic examination of surgical specimen (n = 17).

of only one observer to contour volumes; the other observer was used just for verification. Another limitation is that only the maximum diameter of the tumor on pathologic examination, measured macroscopically in 3 dimensions, was compared with the maximum diameter of the GTV in 3 dimensions. In fact, there are still some other problems, such as tumor size reduction due to formalin fixation before pathologic examination and the interobserver variation in PET/CT target volume delineation. However, our data contribute to the existing body of literature that has attempted to define the target of NSCLC on PET images for radiation therapy planning. It is necessary to compare the volumes on pathologic examination with the GTVs.

CT is the current standard for NSCLC GTV delineation, but despite excellent spatial resolution, substantial variation exists (32-34). According to the available literature, CT scan-based target volume delineation may over- or underestimate the extent of the GTV. Chan et al. (35) found that tumors were larger as measured on CT than as measured on the pathology slice. Therefore, with ¹⁸F-FDG PET, the target volume can either be enlarged, incorporating additional tumor tissue not identified by CT, or reduced, excluding nontumor structures and leading to a reduction of the irradiated normal tissue and, thus, permitting an escalation of the irradiation dose (36). There is no reliable correlation of CT window and level settings to PET SUV, although Hong et al. (37) developed a method of correlating SUV to window and level thresholds. The use of variable CT-threshold settings (18,36,38) can undoubtedly affect the GTV. Caldwell et al. (39) found a reduction in the ratio of largest to smallest GTV by comparing PET/CT coregistered data with CT alone in a group of 30 patients. Although CT and PET may define a quantitatively similar volume, substantial variations may occur in the qualitative definition of the target, with PET offering valuable metabolic information that could result in enlargements or reductions in the target size. In our previous data, the significance of the change in volume from the CT-based target varied greatly (14). CT window and level settings of 1,600 and -300 HU, respectively, for the primary tumor provided the best correlation with the maximum pathologic diameter (14). Therefore, in this study, we chose these settings for CT contouring.

CONCLUSION

Our study compared automatic contouring and manual contouring in delineation of GTV for NSCLC. Contouring using a manual method correlated better with maximum diameter of the primary tumor on pathologic examination and provided a better matching index in the composite PET/CT GTV. In the PET/CT-delineated target volumes for the primary lung tumor, CT had a good matching index and should still be the basis of the defined GTV. When using a 50% intensity level to contour the target of NSCLC, one should consider manual contouring of ¹⁸F-FDG PET to check for any missed disease.

ACKNOWLEDGMENTS

This study was supported by the Ontario Cancer Research Network (OCRN). Results from this study have been presented at the ASTRO annual meeting in Boston, September 21–24, 2008, and at the CARO annual scientific meeting in Montreal, Canada, September 10–13, 2008.

REFERENCES

- Ung YC, Maziak DE, Vanderveen JA, et al. 18 fluorodeoxyglucose positron emission tomography in the diagnosis and staging of lung cancer: a systematic review. J Natl Cancer Inst. 2007;99:1753–1767.
- Ashamalla H, Rafla S, Parikh K, et al. The contribution of integrated PET/CT to the evolving definition of treatment volumes in radiation treatment planning in lung cancer. *Int J Radiat Oncol Biol Phys.* 2005;63:1016–1023.
- Black QC, Grills IS, Kestin LL, et al. Defining a radiotherapy target with positron emission tomography. Int J Radiat Oncol Biol Phys. 2004;60:1272–1282.
- Erdi YE, Rosenzweig K, Erdi AK, et al. Radiotherapy treatment planning for patients with non-small cell lung cancer using positron emission tomography (PET). *Radiother Oncol.* 2002;62:51–60.
- Mah K, Caldwell CB, Ung YC, et al. The impact of ¹⁸FDG-PET on target and critical organs in CT-based treatment planning of patients with poorly defined non-small-cell lung carcinoma: a prospective study. *Int J Radiat Oncol Biol Phys.* 2002;52:339–350.
- Deniaud-Alexandre E, Touboul E, Lerouge D, et al. Impact of computed tomography and ¹⁸F-deoxyglucose coincidence detection emission tomography image fusion for optimization of conformal radiotherapy in non-small-cell lung cancer. *Int J Radiat Oncol Biol Phys.* 2005;63:1432–1441.
- Bradley JD, Ieumwananonthachai N, Purdy JA, et al. Gross tumor volume, critical prognostic factor in patients treated with three-dimensional conformal radiation therapy for non-small-cell lung carcinoma. *Int J Radiat Oncol Biol Phys.* 2002;52:49–57.
- De Ruysscher D, Wanders S, Minken A, et al. Effects of radiotherapy planning with a dedicated combined PET-CT-simulator of patients with non-small cell lung cancer on dose limiting normal tissues and radiation dose-escalation: a planning study. *Radiother Oncol.* 2005;77:5–10.
- Nestle U, Walter K, Schmidt S, et al. ¹⁸F-deoxyglucose positron emission tomography (FDG-PET) for the planning of radiotherapy in lung cancer: high impact in patients with atelectasis. *Int J Radiat Oncol Biol Phys.* 1999;44:593–597.
- Ciernik IF, Dizendorf E, Baumert BG, et al. Radiation treatment planning with an integrated positron emission and computer tomography (PET/CT): a feasibility study. *Int J Radiat Oncol Biol Phys.* 2003;57:853–863.
- van Der Wel A, Nijsten S, Hochstenbag M, et al. Increased therapeutic ratio by ¹⁸FDG-PET CT planning in patients with clinical CT stage N2-N3M0 non-smallcell lung cancer: a modeling study. *Int J Radiat Oncol Biol Phys.* 2005;61:649– 655.
- Kalff V, Hicks RJ, MacManus MP, et al. Clinical impact of ¹⁸F fluorodeoxyglucose positron emission tomography in patients with non-small-cell lung cancer: a prospective study. *J Clin Oncol.* 2001;19:111–118.
- Nestle U, Kremp S, Schaefer-Schuler A, et al. Comparison of different methods for delineation of ¹⁸F-FDG PET-positive tissue for target volume definition in radiotherapy of patients with non-small cell lung cancer. *J Nucl Med.* 2005;46: 1342–1348.
- Wu K, Ung YC, Hornby J, et al. PET CT thresholds for radiotheraphy target definition in non-small-cell lung cancer: how close are we to the pathologic findings? *Int J Radiat Oncol Biol Phys.* 2010;77:699–706.
- Grills IS, Yan D, Black QC, Wong CY, Martinez AA, Kestin LL. Clinical implications of defining the gross tumor volume with combination of CT and ¹⁸FDG-positron emission tomography in non-small-cell lung cancer. *Int J Radiat Oncol Biol Phys.* 2007;67:709–719.
- Nestle U, Kremp S, Grosu A-L. Practical integration of [¹⁸F]-FDG-PET and PET-CT in the planning of radiotherapy for non-small cell lung cancer (NSCLC): the technical basis, ICRU-target volumes, problems, perspectives. *Radiother Oncol.* 2006;81:209–225.
- MacManus M, Nestle U, Rosenzweig KE, et al. Use of PET and PET/CT for radiation therapy planning: IAEA expert report 2006–2007. *Radiother Oncol.* 2009;91:85–94.
- van Baardwijk A, Bosmans G, Boersma L, et al. PET-CT-based auto-contouring in non-small-cell lung cancer correlates with pathology and reduces interobserver variability in the delineation of the primary tumor and involved nodal volumes. *Int J Radiat Oncol Biol Phys.* 2007;68:771–778.

1522 The JOURNAL OF NUCLEAR MEDICINE • Vol. 51 • No. 10 • October 2010

jnm077974-sn ■ 9/9/10

- Kiffer JD, Berlangieri SU, Scott AM, et al. The contribution of ¹⁸F-fluoro-2deoxy-glucose positron emission tomographic imaging to radiotherapy planning in lung cancer. *Lung Cancer*. 1998;19:167–177.
- Nestle U, Hellwig D, Schmidt S, et al. 2-Deoxy-2-[¹⁸F]fluoro-D-glucose positron emission tomography in target volume definition for radiotherapy of patients with non-small-cell lung cancer. *Mol Imaging Biol.* 2002;4:257–263.
- Bradley J, Thorstad WL, Mutic S, et al. Impact of ¹⁸FDG-PET on radiation therapy volume delineation in non-small cell lung cancer. Int J Radiat Oncol Biol Phys. 2004;59:78–86.
- 22. Vanuytsel LJ, Vansteenkiste JF, Stroobants SG, et al. The impact of ¹⁸F-fluoro-2deoxy-D-glucose positron emission tomography (FDG-PET) lymph node staging on the radiation treatment volumes in patients with non-small cell lung cancer. *Radiother Oncol.* 2000;55:317–324.
- Giraud P, Grahek D, Montravers F, et al. CT and ¹⁸F-deoxyglucose (FDG) image fusion for optimization of conformal radiotherapy of lung cancers. *Int J Radiat Oncol Biol Phys.* 2001;49:1249–1257.
- Brianzoni E, Rossi G, Ancidei S, et al. Radiotherapy planning: PET/CT scanner performances in the definition of gross tumour volume and clinical target volume. *Eur J Nucl Med Mol Imaging*. 2005;32:1392–1399.
- Scarfone C, Lavely WC, Cmelak AJ, et al. Prospective feasibility trial of radiotherapy target definition for head and neck cancer using 3-dimensional PET and CT imaging. J Nucl Med. 2004;45:543–552.
- Miller TR, Grigsby PW. Measurement of tumor volume by PET to evaluate prognosis in patients with advanced cervical cancer treated by radiation therapy. *Int J Radiat Oncol Biol Phys.* 2002;53:353–359.
- Giraud P, Antoine M, Larrouy A, et al. Evaluation of microscopic tumor extension in non-small-cell lung cancer for three-dimensional conformal radiotherapy planning. *Int J Radiat Oncol Biol Phys.* 2000;48:1015–1024.
- Erdi YE, Nehmeh SA, Pan T, et al. The CT motion quantitation of lung lesions and its impact on PET-measured SUVs. J Nucl Med. 2004;45:1287–1292.
- Clarke GM, Eidt S, Sun L, Mawdsley G, Zubovits JT, Yaffe MJ. Whole-specimen histopathology: a method to produce whole-mount breast serial sections for 3-D digital histopathology imaging. *Histopathology*. 2007;50:232–242.

- Stroom J, Blaauwgeers H, van Baardwijk A, et al. Feasibility of pathologycorrelated lung imaging for accurate target definition of lung tumors. *Int J Radiat Oncol Biol Phys.* 2007;69:267–275.
- Geets X, Daisne JF, Arcangeli S, et al. Inter-observer variability in the delineation of pharyngo-laryngeal tumor, parotid glands and cervical spinal cord: comparison between CT-scan and MRI. *Radiother Oncol.* 2005;77: 25–31.
- Steenbakkers RJ, Duppen JC, Fitton I, et al. Observer variation in target volume delineation of lung cancer related to radiation oncologist-computer interaction: a 'Big Brother' evaluation. *Radiother Oncol.* 2005;77:182–190.
- Caldwell CB, Mah K, Skinner M, Danjoux CE. Can PET provide the 3D extent of tumor motion for individualized internal target volumes? A phantom study of the limitations of CT and the promise of PET. *Int J Radiat Oncol Biol Phys.* 2003;55:1381–1393.
- Dahele MR, Ung YC. ¹⁸F-FDG PET in planning radiation treatment of non-small cell lung cancer: where exactly is the tumor? [letter]. *J Nucl Med.* 2007; 48:1402.
- Chan R, He Y, Haque A, Zwischenberger J. Computed tomographic-pathologic correlation of gross tumor volume and clinical target volume in non-small cell lung cancer: a pilot experience. *Arch Pathol Lab Med.* 2001;125:1469–1472.
- Jennings SG, Winer-Muram HT, Tann M, Ying J, Dowdeswell I. Distribution of stage I lung cancer growth rates determined with serial volumetric CT measurements. *Radiology*. 2006;241:554–563.
- Hong R, Halama J, Bova D, Sethi A, Emami B. Correlation of PET standard uptake value and CT window-level thresholds for target delineation in CT-based radiation treatment planning. *Int J Radiat Oncol Biol Phys.* 2007;67:720–726.
- Saito H, Minamiya Y, Kawai H, et al. Usefulness of circumference difference for estimating the likelihood of malignancy in small solitary pulmonary nodules on CT. *Lung Cancer.* 2007;58:348–354.
- Caldwell CB, Mah K, Ung YC, et al. Observer variation in contouring gross tumor volume in patients with poorly defined non-small-cell lung tumors on CT: the impact of ¹⁸FDG-hybrid PET fusion. *Int J Radiat Oncol Biol Phys.* 2001;51: 923–931.